

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 5171  
CALIBRATION DATE: 18-Feb-10

SBE3 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## ITS-90 COEFFICIENTS

g = 4.39251665e-003  
h = 6.45697898e-004  
i = 2.31743619e-005  
j = 2.18386423e-006  
f0 = 1000.0

## IPTS-68 COEFFICIENTS

a = 3.68121226e-003  
b = 6.01411298e-004  
c = 1.57101147e-005  
d = 2.18537559e-006  
f0 = 3137.684

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	3137.684	-1.5000	0.00000
1.0000	3318.114	1.0000	-0.00001
4.5000	3583.149	4.5000	0.00003
8.0000	3863.044	8.0000	-0.00003
11.5000	4158.208	11.5000	-0.00001
15.0000	4469.021	15.0000	0.00003
18.5000	4795.850	18.5000	0.00003
22.0000	5139.050	21.9999	-0.00006
25.5000	5498.987	25.5000	-0.00002
29.0000	5875.986	29.0001	0.00006
32.5000	6270.343	32.5000	-0.00002

Temperature ITS-90 =  $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$  (°C)

Temperature IPTS-68 =  $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$  (°C)

Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35 °C)

Residual = instrument temperature - bath temperature

