

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0730
CALIBRATION DATE: 26-Aug-04p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.3355

Boc = 0.0000

Voffset = -0.5005

TCor = 0.0009

PCor = 1.350e-04

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.39	12.00	0.01	1.045	1.39	0.00
1.40	20.00	0.01	1.147	1.40	-0.00
1.44	6.00	0.00	0.990	1.44	0.00
1.47	30.00	0.01	1.309	1.47	0.00
1.48	26.00	0.01	1.259	1.48	0.00
1.48	2.00	0.00	0.957	1.48	-0.00
4.12	12.02	0.01	2.114	4.12	0.00
4.16	20.00	0.01	2.410	4.15	-0.01
4.23	26.00	0.01	2.671	4.23	-0.00
4.25	6.00	0.00	1.950	4.26	0.01
4.29	30.00	0.01	2.858	4.29	0.00
4.38	2.00	0.00	1.846	4.37	-0.01
7.02	30.00	0.01	4.364	7.03	0.01
7.14	12.04	0.01	3.304	7.16	0.02
7.27	6.00	0.00	2.983	7.29	0.01
7.39	2.00	0.00	2.769	7.37	-0.02
7.40	20.00	0.01	3.903	7.39	-0.01
7.48	26.00	0.01	4.337	7.47	-0.01

oxygen (ml/l) = (Soc * (V + Voffset)) * exp(Tcor * T) * Oxsat(T,S) * exp(PCor * P)

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU]

Oxsat(T,S) = oxygen saturation [ml/l], P = pressure [dbar]

Residual = instrument oxygen - bath oxygen

