

Conductivity Calibration Report

Customer:	US Department of Commerce, NOAA					
Job Number:	67586	Date of Report:	1/26/2012			
Model Number:	SBE 04C	Serial Number:	043647			

Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.

'AS RECEIVED CALIBRATION'	✓ Perfor			Not Performed	
Date: 1/25/2012	Drift since last cal:	0.0	000] PSU/month*	
Comments:					

'CALIBR	RATION A	FTER CLEANING & REPLATINIZING'	Perf	ormed	Not	Performed
Date:		Drift since I	Last cal:			PSU/month*

Comments:

*Measured at 3.0 S/m

Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.