

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1701
CALIBRATION DATE: 22-Jun-12

SBE3 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.79080984e-003
h = 6.54354306e-004
i = 1.87219571e-005
j = 1.04636291e-006
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121221e-003
b = 5.97871152e-004
c = 1.31700294e-005
d = 1.04751938e-006
f0 = 5912.710

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	5912.710	-1.5000	0.00001
1.0000	6254.701	0.9999	-0.00007
4.5000	6757.024	4.5001	0.00009
8.0000	7287.425	8.0000	0.00002
11.5000	7846.701	11.5000	-0.00001
15.0000	8435.599	15.0000	0.00004
18.5000	9054.785	18.4998	-0.00017
22.0000	9705.060	21.9999	-0.00006
25.5000	10387.070	25.5001	0.00012
29.0000	11101.425	29.0001	0.00015
32.5000	11848.732	32.4999	-0.00012

Temperature ITS-90 = $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)

Temperature IPTS-68 = $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature

