The sensor was backfilled with Argon gas.

Temperature Calibration Report

Customer:	US Department of Commerce, NOAA			
Job Number:	69711	Date of Report:		7/10/2012
Model Number	SBE 03-02/F	Serial Number:		031701
Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request. An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users				
must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data. 'AS RECEIVED CALIBRATION' Performed Not Performed				
Date: 6/22/2012		Drift since last cal:	0.00099	Degrees Celsius/year
Comments:	_			
'FINAL CALIBRA Date: 7/10/2012		✓ Perform Drift since 12 Mar 10	ned -0.00085	Not Performed Degrees Celsius/year
Comments:				