

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1609  
CALIBRATION DATE: 14-May-13

SBE3 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## ITS-90 COEFFICIENTS

g = 4.86616718e-003  
h = 6.80226179e-004  
i = 2.63390016e-005  
j = 2.04340646e-006  
f0 = 1000.0

## IPTS-68 COEFFICIENTS

a = 3.68121025e-003  
b = 6.03712862e-004  
c = 1.49874264e-005  
d = 2.04485119e-006  
f0 = 6399.456

| BATH TEMP<br>(ITS-90) | INSTRUMENT FREQ<br>(Hz) | INST TEMP<br>(ITS-90) | RESIDUAL<br>(ITS-90) |
|-----------------------|-------------------------|-----------------------|----------------------|
| -1.5000               | 6399.456                | -1.4998               | 0.00015              |
| 1.0000                | 6765.946                | 0.9999                | -0.00009             |
| 4.5000                | 7304.066                | 4.4998                | -0.00023             |
| 8.0000                | 7872.115                | 7.9999                | -0.00014             |
| 11.5000               | 8470.879                | 11.5002               | 0.00020              |
| 15.0000               | 9101.071                | 15.0005               | 0.00055              |
| 18.5000               | 9763.196                | 18.4998               | -0.00023             |
| 22.0000               | 10458.309               | 21.9998               | -0.00024             |
| 25.5000               | 11186.975               | 25.4999               | -0.00008             |
| 29.0000               | 11949.799               | 29.0000               | 0.00004              |
| 32.4999               | 12747.363               | 32.5000               | 0.00007              |

Temperature ITS-90 =  $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$  (°C)

Temperature IPTS-68 =  $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$  (°C)

Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35 °C)

Residual = instrument temperature - bath temperature

