



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA

Website: <http://www.seabird.com>

Phone: (425) 643-9866

FAX: (425) 643-9954

Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 11/13/2012 Description SBE-4 Conductivity Sensor

Job Number: stock

Customer Name

SBE Sensor Information:

Model Number: 04

Serial Number: 4222

Pressure Sensor Information:

Sensor Type: None

Sensor Serial Number: None

Sensor Rating: 0

Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test: ☒

Tested By: VG

