



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA

Website: <http://www.seabird.com>

Phone: (425) 643-9866

FAX: (425) 643-9954

Email: [seabird@seabird.com](mailto:seabird@seabird.com)

## SBE Pressure Test Certificate

Test Date: 7/1/2013 Description SBE-43 DO Sensor

Job Number: Stock Customer Name

### SBE Sensor Information:

Model Number: 43

Serial Number: 2691

### Pressure Sensor Information:

Sensor Type:

Sensor Serial Number:

Sensor Rating: 0

### Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 30 Minutes

Passed Test: ☒

Tested By: RH

