



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA

Website: <http://www.seabird.com>

Phone: (425) 643-9866

FAX: (425) 643-9954

Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 10/22/2014 Description SBE-5T Submersible Pump

SBE Sensor Information:

Model Number: 5T

Serial Number: 7890

Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test: ☒

Tested By: AP

