



## Sea-Bird Electronics, Inc.

13431 NE 20<sup>th</sup> St. Bellevue, Washington 98005 USA  
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

## Pressure Test Certificate

Test Date: 2017-01-26

Description: SBE-4 Conductivity Sensor

### Sensor Information:

Model Number: SBE-4

Serial Number: 4709

### Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For: 15 Minutes

High Pressure Test: 10000 PSI Held For: 15 Minutes

Passed Test: True

Tested By: sa

High pressure is generally equal to the maximum depth rating of the instrument

