



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2017-07-27**

Description: **SBE-5T Submersible Pump**

### Sensor Information:

Model Number: **SBE-5T**

Serial Number: **9256**

### Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **10000** PSI Held For: **15** Minutes

Passed Test: **True**

Tested By: **AM**

High pressure is generally equal to the maximum depth rating of the instrument

