



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866231

CUSTOMER INFORMATION

Name: NOAA - ATLANTIC OCEANOGRAPHIC AND METEOROLOGICAL LAB
 Account : 40279759
 GRANT RAWSON
 GRANT.RAWSON@NOAA.GOV
 (305) 361-4363

PO Number:

Bill To Address

4301 RICKENBACKER CAUSEWAY
 MIAMI,FL,33149,US

Ship To Address

4301 RICKENBACKER CAUSEWAY
 MIAMI,FL,33149,US

PRODUCT INFORMATION

Item: 43.LEGACY
Item Description: (LEGACY) SBE 43 Oxygen Sensor
Serial: 432085

Special Notes

Services Requested:
 Dissolved Oxygen Calibration

No problems found.

Services Performed:
 Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866231

PRODUCT INFORMATION

Item: 43.LEGACY
Item Description: (LEGACY) SBE 43 Oxygen Sensor
Serial: 43-2949

Special Notes

Services Requested:
 Dissolved Oxygen Calibration

No problems found.

Services Performed:
 Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866231

PRODUCT INFORMATION

Item: 43.LEGACY
Item Description: (LEGACY) SBE 43 Oxygen Sensor
Serial: 430730

Special Notes

Services Requested:
 Dissolved Oxygen Calibration

No problems found.

Services Performed:
 Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866231

PRODUCT INFORMATION

Item: 04C.LEGACY
Item Description: (LEGACY) SBE 04C Conductivity Sensor
Serial: 042980

Special Notes

Services Requested:
 Conductivity Calibration

No problems found.

Services Performed:
 Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866231

PRODUCT INFORMATION

Item: 04C.LEGACY
Item Description: (LEGACY) SBE 04C Conductivity Sensor
Serial: 042973

Special Notes
 Services Requested:
 Conductivity Calibration

No problems found.

Services Performed:
 Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty