



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: **2020-03-23**

Description: **SBE-3 Temperature Sensor**

Sensor Information:

Replaced the main piston "O"-Rings.
Replaced temperature probe.

Model Number: **SBE-3**

Serial Number: **5233**

Pressure Test Protocol:

Low Pressure Test: **50** PSI Held For: **15** Minutes

High Pressure Test: **10000** PSI Held For: **30** Minutes

Passed Test: **True**

Tested By: **BLT**

