



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0353
CALIBRATION DATE: 25-Jan-22

Glider APL CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.97122842e+000
h = 1.11843624e+000
i = -1.64825012e-003
j = 2.10040559e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.98993	0.00000	0.00000
1.0162	34.6494	2.96445	5.95796	2.96462	0.00016
4.5000	34.6292	3.26881	6.18178	3.26861	-0.00020
15.0000	34.5875	4.24650	6.85167	4.24648	-0.00002
18.5000	34.5786	4.59023	7.07186	4.59027	0.00004
24.0000	34.5687	5.14585	7.41364	5.14592	0.00007
29.0000	34.5627	5.66544	7.71918	5.66539	-0.00005
32.5000	34.5587	6.03613	7.92975	6.03594	-0.00019

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

