



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1387
CALIBRATION DATE: 21-Apr-22

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.06572082e+001
h = 1.59574146e+000
i = -8.31072724e-004
j = 1.53892127e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.58519	0.00000	0.00000
-1.0001	34.7274	2.79807	4.92123	2.79806	-0.00001
0.9999	34.7274	2.96908	5.02887	2.96909	0.00001
14.9999	34.7263	4.26173	5.77752	4.26173	0.00000
18.4999	34.7239	4.60742	5.96172	4.60743	0.00000
28.9999	34.7173	5.68792	6.50360	5.68790	-0.00002
32.5000	34.7032	6.05849	6.67926	6.05851	0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

