



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0266  
 CALIBRATION DATE: 31-Jan-19

Glider APL CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.82066234e+000  
 h = 1.15571339e+000  
 i = -1.69732895e-003  
 j = 2.01388758e-004

CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.91914	0.00000	0.00000
1.0000	34.9011	2.98252	5.86469	2.98253	0.00001
4.5000	34.8798	3.29012	6.08754	3.29011	-0.00001
15.0000	34.8350	4.27366	6.75039	4.27366	-0.00001
18.5000	34.8250	4.61940	6.96822	4.61940	-0.00000
24.0000	34.8132	5.17822	7.30635	5.17823	0.00002
29.0000	34.8026	5.70033	7.60841	5.70032	-0.00001
32.5000	34.7927	6.07234	7.81635	6.07228	-0.00006

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

