



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0266  
 CALIBRATION DATE: 29-Jan-21

Glider APL CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.82631203e+000  
 h = 1.15803127e+000  
 i = -2.40831796e-003  
 j = 2.57588079e-004

CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.91906	0.00000	0.00000
0.9999	34.6543	2.96343	5.85078	2.96343	0.00001
4.4999	34.6345	3.26925	6.07290	3.26925	-0.00000
15.0000	34.5934	4.24715	6.73351	4.24712	-0.00003
18.4999	34.5848	4.59095	6.95064	4.59096	0.00001
23.9999	34.5756	5.14675	7.28770	5.14679	0.00003
29.0000	34.5707	5.66660	7.58907	5.66658	-0.00002
32.5000	34.5690	6.03772	7.79682	6.03749	-0.00023

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

