



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0313  
 CALIBRATION DATE: 09-Feb-21

Glider APL CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.01077809e+001  
 h = 1.09928025e+000  
 i = -2.92953389e-003  
 j = 2.80618185e-004

CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	3.04106	0.00000	0.00000
1.0000	34.6683	2.96452	6.03407	2.96453	0.00001
4.4999	34.6489	3.27047	6.26164	3.27047	-0.00001
15.0000	34.6084	4.24880	6.93873	4.24878	-0.00002
18.5000	34.5997	4.59273	7.16130	4.59273	0.00000
24.0000	34.5906	5.14875	7.50686	5.14878	0.00003
29.0000	34.5863	5.66887	7.81590	5.66886	-0.00001
32.5000	34.5838	6.04002	8.02893	6.03995	-0.00007

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

