



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0313
 CALIBRATION DATE: 25-Jan-22

Glider APL CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00767141e+001 CPcor = -9.5700e-008 (nominal)
 h = 1.09154374e+000 CTcor = 3.2500e-006 (nominal)
 i = -1.13425190e-003
 j = 1.68919067e-004

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	3.04099	0.00000	0.00000
1.0162	34.6494	2.96445	6.03439	2.96452	0.00007
4.5000	34.6292	3.26881	6.26061	3.26872	-0.00008
15.0000	34.5875	4.24650	6.93725	4.24648	-0.00002
18.5000	34.5786	4.59023	7.15973	4.59025	0.00002
24.0000	34.5687	5.14585	7.50512	5.14589	0.00004
29.0000	34.5627	5.66544	7.81397	5.66541	-0.00003
32.5000	34.5587	6.03613	8.02684	6.03599	-0.00014

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

