



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0349  
 CALIBRATION DATE: 11-Feb-21

Glider APL CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00861846e+001  
 h = 1.14623715e+000  
 i = -2.81015966e-003  
 j = 2.86493531e-004

CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.97394	0.00000	0.00000
1.0000	34.6750	2.96504	5.90491	2.96504	0.00001
4.5000	34.6553	3.27103	6.12769	3.27103	-0.00000
15.0000	34.6148	4.24950	6.79049	4.24948	-0.00002
18.5000	34.6062	4.59350	7.00839	4.59351	0.00002
24.0000	34.5973	5.14964	7.34666	5.14964	0.00000
28.9999	34.5925	5.66976	7.64917	5.66976	-0.00000
32.5001	34.5728	6.03832	7.85770	6.04089	0.00257

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

