



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2085
CALIBRATION DATE: 01-Jan-20

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS:
Soc = 0.4457
Voffset = -0.4896
Tau20 = 1.37
A = -3.8245e-003
B = 2.0359e-004
C = -3.0039e-006
E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS
D1 = 1.92634e-4
D2 = -4.64803e-2
H1 = -3.300000e-2
H2 = 5.00000e+3
H3 = 1.45000e+3

BATH OXYGEN (ml/l)	BATH TEMPERATURE (° C)	BATH SALINITY (PSU)	INSTRUMENT OUTPUT (volts)	INSTRUMENT OXYGEN (ml/l)	RESIDUAL (ml/l)
1.17	20.00	0.00	0.911	1.17	0.00
1.18	30.00	0.00	0.995	1.18	0.00
1.18	2.00	0.00	0.763	1.17	-0.00
1.18	26.00	0.00	0.962	1.18	0.00
1.18	12.00	0.00	0.848	1.18	-0.00
1.19	6.00	0.00	0.802	1.19	0.00
3.93	26.00	0.00	2.069	3.94	0.01
3.93	30.00	0.00	2.180	3.93	0.00
3.95	2.00	0.00	1.409	3.94	-0.01
3.96	20.00	0.00	1.912	3.96	-0.00
3.97	6.00	0.00	1.532	3.98	0.01
3.98	12.00	0.00	1.698	3.97	-0.01
6.73	2.00	0.00	2.062	6.73	0.00
6.79	30.00	0.00	3.406	6.79	-0.00
6.81	26.00	0.00	3.224	6.82	0.01
6.83	12.02	0.00	2.568	6.83	0.00
6.84	6.00	0.00	2.280	6.84	-0.00
6.87	20.00	0.00	2.955	6.86	-0.01

V = instrument output (volts); T = temperature (°C); S = salinity (PSU); K = temperature (°K)

Oxsol(T,S) = oxygen saturation (ml/l); P = pressure (dbar)

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * Oxsol(T,S) * exp(E * P / K)

Residual (ml/l) = instrument oxygen - bath oxygen

