

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1346
CALIBRATION DATE: 03-Jul-12

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.06004422e+000
h = 5.34692909e-001
i = 4.28358363e-005
j = 2.94920362e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 4.11176894e-005
b = 5.34790299e-001
c = -4.06038160e+000
d = -8.49380846e-005
m = 3.9
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.75470	0.00000	0.00000
-0.9986	34.6588	2.79319	7.72008	2.79320	0.00002
1.0000	34.6593	2.96382	7.92253	2.96381	-0.00001
15.0000	34.6606	4.25453	9.31058	4.25450	-0.00003
18.5000	34.6605	4.59993	9.64769	4.59993	-0.00000
29.0000	34.6584	5.67936	10.63113	5.67943	0.00007
32.5001	34.6513	6.05047	10.94827	6.05043	-0.00005

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

