

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1347
CALIBRATION DATE: 24-Aug-10

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.08968741e+000
h = 5.37539075e-001
i = -2.12874748e-005
j = 3.20735069e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 3.04299825e-005
b = 5.37471314e-001
c = -4.08953948e+000
d = -8.40875707e-005
m = 4.0
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.75781	0.00000	0.00000
-1.0000	34.8168	2.80461	7.71943	2.80462	0.00001
1.0000	34.8171	2.97603	7.92193	2.97603	0.00000
15.0000	34.8186	4.27187	9.30944	4.27182	-0.00005
18.5000	34.8181	4.61858	9.64642	4.61860	0.00002
29.0000	34.8170	5.70242	10.62963	5.70248	0.00006
32.5000	34.8107	6.07512	10.94677	6.07508	-0.00004

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

