

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1347
CALIBRATION DATE: 22-Jun-12

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.08938570e+000
h = 5.37556494e-001
i = -1.86553989e-005
j = 3.19234096e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 3.04125853e-005
b = 5.37502367e-001
c = -4.08931096e+000
d = -8.44280442e-005
m = 4.0
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.75765	0.00000	0.00000
-1.0000	34.7336	2.79853	7.71193	2.79856	0.00003
1.0000	34.7345	2.96964	7.91421	2.96962	-0.00001
15.0000	34.7353	4.26273	9.30011	4.26267	-0.00005
18.5000	34.7348	4.60872	9.63672	4.60874	0.00002
29.0000	34.7324	5.69012	10.61871	5.69020	0.00008
32.5000	34.7258	6.06199	10.93544	6.06194	-0.00006


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



24-Aug-10 0.9999362
22-Jun-12 1.0000000