

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2980
CALIBRATION DATE: 15-Feb-11

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.00397788e+001
h = 1.37167095e+000
i = -5.28485306e-005
j = 7.39871538e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 8.47207596e-005
b = 1.37143773e+000
c = -1.00392532e+001
d = -8.35269924e-005
m = 3.9
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.70504	0.00000	0.00000
-1.0000	34.9805	2.81657	5.27420	2.81658	0.00001
1.0000	34.9810	2.98870	5.39152	2.98868	-0.00001
15.0000	34.9802	4.28958	6.20667	4.28958	-0.00001
18.5000	34.9797	4.63770	6.40713	4.63771	0.00001
29.0000	34.9790	5.72596	6.99654	5.72595	-0.00001
32.5001	34.9723	6.10012	7.18795	6.10012	0.00000

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

