

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2980
CALIBRATION DATE: 15-Oct-14

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.01280741e+001
h = 1.38228305e+000
i = 6.28668686e-004
j = 2.77252284e-005

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.70499	0.00000	0.00000
-1.0000	34.6407	2.79175	5.23863	2.79177	0.00002
1.0000	34.6418	2.96247	5.35466	2.96245	-0.00001
15.0000	34.6427	4.25256	6.16099	4.25253	-0.00003
18.5000	34.6424	4.59778	6.35937	4.59780	0.00002
29.0000	34.6406	5.67677	6.94266	5.67681	0.00004
32.5001	34.6340	6.04780	7.13210	6.04777	-0.00003

f = INST FREQ / 1000.0

Conductivity = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$ Siemens / meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity

