



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2980
CALIBRATION DATE: 04-Oct-18

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.01328733e+001
h = 1.38372060e+000
i = 2.68719561e-004
j = 5.51784825e-005

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.70498	0.00000	0.00000
-1.0000	34.5398	2.78437	5.23329	2.78435	-0.00002
0.9997	34.5410	2.95464	5.34918	2.95467	0.00003
15.0000	34.5370	4.24096	6.15392	4.24094	-0.00002
18.5000	34.5343	4.58498	6.35180	4.58497	-0.00000
29.0001	34.5276	5.66034	6.93374	5.66038	0.00004
32.5001	34.5144	6.02928	7.12233	6.02926	-0.00002

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

