



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 3833  
CALIBRATION DATE: 12-Dec-18

SBE 4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.04852317e+001  
h = 1.37172365e+000  
i = -7.72476647e-004  
j = 1.26498540e-004

CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

| BATH TEMP<br>(° C) | BATH SAL<br>(PSU) | BATH COND<br>(S/m) | INSTRUMENT<br>OUTPUT (kHz) | INSTRUMENT<br>COND (S/m) | RESIDUAL<br>(S/m) |
|--------------------|-------------------|--------------------|----------------------------|--------------------------|-------------------|
| 0.0000             | 0.0000            | 0.00000            | 2.76593                    | 0.00000                  | 0.00000           |
| -1.0000            | 34.6346           | 2.79130            | 5.29185                    | 2.79129                  | -0.00001          |
| 1.0000             | 34.6349           | 2.96193            | 5.40805                    | 2.96194                  | 0.00001           |
| 15.0000            | 34.6345           | 4.25166            | 6.21594                    | 4.25166                  | -0.00001          |
| 18.5000            | 34.6328           | 4.59665            | 6.41471                    | 4.59666                  | 0.00002           |
| 29.0000            | 34.6274           | 5.67485            | 6.99925                    | 5.67483                  | -0.00003          |
| 32.5000            | 34.6173           | 6.04520            | 7.18899                    | 6.04522                  | 0.00002           |

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Conductivity (S/m) =  $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

