



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3860
CALIBRATION DATE: 08-Oct-19

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.03264156e+001
h = 1.48466765e+000
i = -1.10703095e-003
j = 1.70253190e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.63885	0.00000	0.00000
-1.0000	34.8080	2.80397	5.08555	2.80396	-0.00001
1.0000	34.8088	2.97538	5.19781	2.97539	0.00001
15.0000	34.8088	4.27079	5.97793	4.27080	0.00001
18.5000	34.8076	4.61734	6.16979	4.61733	-0.00001
29.0000	34.7985	5.69973	6.73361	5.69973	0.00000
32.5001	34.7820	6.07070	6.91613	6.07063	-0.00007

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

