



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3860
CALIBRATION DATE: 23-Jun-21

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.03177755e+001
h = 1.48171737e+000
i = -3.17540012e-004
j = 1.17016526e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.63884	0.00000	0.00000
-1.0000	34.8207	2.80490	5.08613	2.80490	0.00000
1.0000	34.8212	2.97634	5.19836	2.97634	-0.00000
15.0000	34.8214	4.27217	5.97849	4.27217	0.00000
18.5000	34.8199	4.61879	6.17033	4.61879	-0.00001
29.0001	34.8130	5.70185	6.73435	5.70186	0.00001
32.5001	34.7990	6.07333	6.91710	6.07332	-0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

