

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4223
CALIBRATION DATE: 30-Mar-16

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.93263698e+000
h = 1.37445308e+000
i = -2.40165079e-003
j = 2.29741004e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.69294	0.00000	0.00000
-1.0000	34.7028	2.79628	5.26286	2.79627	-0.00001
1.0000	34.7031	2.96721	5.38025	2.96722	0.00001
15.0000	34.7031	4.25919	6.19562	4.25919	-0.00000
18.5000	34.7025	4.60490	6.39608	4.60493	0.00003
29.0001	34.7016	5.68565	6.98530	5.68560	-0.00005
32.5001	34.6922	6.05680	7.17642	6.05684	0.00003

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

