



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 6379
 CALIBRATION DATE: 29-Feb-24

SBE 4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.01690880e+001
 h = 1.52076197e+000
 i = -1.20294686e-003
 j = 1.76788457e-004

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.58753	0.00000	0.00000
-1.0001	34.5080	2.78204	5.00067	2.78204	0.00001
0.9999	34.5090	2.95218	5.11132	2.95216	-0.00002
15.0000	34.5094	4.23793	5.88026	4.23797	0.00004
18.5000	34.5094	4.58203	6.06938	4.58200	-0.00002
29.0000	34.5037	5.65685	6.62524	5.65683	-0.00002
32.5000	34.4889	6.02532	6.80525	6.02534	0.00002

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

