



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

Pressure Test Certificate

Test Date: **8/1/2014**

Description: **SBE-5T Submersible Pump**

Sensor Information:

Model Number: **5T**

none

Serial Number: **7741**

15000

Pressure Test Protocol:

Low Pressure Test: **40** PSI

Held For: **15**

Minutes

High Pressure Test: **10000** PSI

Held For: **15**

Minutes

Passed Test: **Yes**

Tested By: **AP**

**High pressure is
generally equal
to the maximum
depth rating of
the instrument**

