

SYSTEM CONFIGURATION

SBE 9plus CTD Underwater Unit	S/N 09-1292
Depth Capability	6800 meters
Pressure Sensor Range	10000 psia
Digiquartz Pressure Sensor (w/ temp-comp)	S/N 136924
Modulo 12P	S/N MOD12P-1304
AD590M (Pressure Temperature Coefficient)	0.01279955
AD590B (Pressure Temperature Coefficient)	-9.15105924
A/D Input Voltage Range	0 to 5 Volts DC

SEASOFT CONFIGURATION:

The settings for the configuration of your instrument as delivered are documented below:

Configuration for the SBE 911plus/917plus CTD

Configuration file opened: 09-1292.xmlcon

Frequency channels suppressed Voltage words suppressed (1 word = 2 channels)

Deck unit or SEARAM

Computer interface

Scans to average

NMEA position data added NMEA depth data added
 NMEA device connected to deck unit NMEA time added
 NMEA device connected to PC

Surface PAR voltage added Scan time added

Channel	Sensor
1. Frequency	Temperature
2. Frequency	Conductivity
3. Frequency	Pressure, Digiquartz with TC

New
Open...
Save
Save As...
Select...
Modify...

Report... Help... Exit Cancel