



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866230

CUSTOMER INFORMATION

Name: NOAA - ATLANTIC OCEANOGRAPHIC AND METEOROLOGICAL LAB
 Account : 40279759
 GRANT RAWSON
 GRANT.RAWSON@NOAA.GOV
 (305) 361-4363

PO Number:

Bill To Address

4301 RICKENBACKER CAUSEWAY
 MIAMI,FL,33149,US

Ship To Address

4301 RICKENBACKER CAUSEWAY
 MIAMI,FL,33149,US

PRODUCT INFORMATION

Item: 43.LEGACY
Item Description: (LEGACY) SBE 43 Oxygen Sensor
Serial: 432712

Special Notes

Services Requested:
 Dissolved Oxygen Calibration

No problems found.

Services Performed:
 Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866230

PRODUCT INFORMATION

Item: 03F.LEGACY
Item Description: (LEGACY) SBE 03F Temperature Sensor
Serial: 031692

Special Notes

Services Requested:
 Temperature Calibration

No problems found.

Services Performed:
 Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_3	SBE 3 Temperature sensor calibration; 11 points -1 to 31deg.C	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866230

PRODUCT INFORMATION

Item: 03P.LEGACY
Item Description: (LEGACY) SBE 03Plus Temperature Sensor
Serial: 03P5140

Special Notes

Services Requested:
 Temperature Calibration

No problems found.

Services Performed:
 Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_3	SBE 3 Temperature sensor calibration; 11 points -1 to 31deg.C	1

Unbilled Items

Item	Item Description	Qty