



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866232

CUSTOMER INFORMATION

Name: NOAA - ATLANTIC OCEANOGRAPHIC AND METEOROLOGICAL LAB
 Account : 40279759
 GRANT RAWSON
 GRANT.RAWSON@NOAA.GOV
 (305) 361-4363

PO Number:

Bill To Address

4301 RICKENBACKER CAUSEWAY
 MIAMI,FL,33149,US

Ship To Address

4301 RICKENBACKER CAUSEWAY
 MIAMI,FL,33149,US

PRODUCT INFORMATION

Item: 03F.LEGACY
Item Description: (LEGACY) SBE 03F Temperature Sensor
Serial: 031075

Special Notes

Services Requested:
 Temperature Calibration

Problems Found:

The sensor anode was found to have reached the end of its life. Replacement required.

Services Performed:

Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.
 Installed NEW anode.

Item	Item Description	Qty
23041	Zinc anode ring	1
CAL_3	SBE 3 Temperature sensor calibration; 11 points -1 to 31deg.C	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866232

PRODUCT INFORMATION

Item: 04C.LEGACY
Item Description: (LEGACY) SBE 04C Conductivity Sensor
Serial: 041374

Special Notes
 Services Requested:
 Conductivity Calibration

Problems Found:
 The sensor anode was found to have reached the end of its life. Replacement required.

Services Performed:
 Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.
 Installed NEW anode.

Item	Item Description	Qty
23548	Zinc anode ring (sensors with tubing connector)	1
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005509738
 24-JAN-2020
 315866232

PRODUCT INFORMATION

Item: 04C.LEGACY
Item Description: (LEGACY) SBE 04C Conductivity Sensor
Serial: 041387

Special Notes

Services Requested:
 Conductivity Calibration

Problems Found:

The sensor anode was found to have reached the end of its life. Replacement required.

Services Performed:

Performed initial diagnostic evaluation.
 Performed "POST" cruise calibration.
 Installed NEW anode.

Item	Item Description	Qty
23548	Zinc anode ring (sensors with tubing connector)	1
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty