



**Sea-Bird Electronics, Inc.**  
 13431 NE 20th Street  
 Bellevue, WA 98005 United States

Phone  
 Fax

+1-425-643-9866  
 +1-425-643-9954  
 www.seabird.com

**SERVICE REPORT**

**Service Request**  
**Date**  
**Sales Order**

**1005510402**  
 22-MAY-2020  
 316003277

**CUSTOMER INFORMATION**

Name: NOAA - ATLANTIC OCEANOGRAPHIC AND METEOROLOGICAL LAB  
 Account : 40279759

PO Number:  
 IDIQ EA133M-15-CQ-0039(TASK  
 1333MJ20FNRMA0011)

GRANT RAWSON  
 GRANT.RAWSON@NOAA.GOV  
 (305) 361-4363

**Bill To Address**  
 4301 RICKENBACKER CAUSEWAY  
 MIAMI,FL,33149,US

**Ship To Address**  
 4301 RICKENBACKER CAUSEWAY  
 MIAMI,FL,33149,US

REQUESTED TRANSACTION DETAILS:  
 PO NAME: Grant Rawson  
 PO NUMBER: Credit Card Payment  
 PAPER COPIES: Yes  
 REQUESTED RETURN: 27-Mar-2020  
 SPLIT SHIP: split

DUE TO THE LOCAL COVID 19 ISSUES. ONLY DIGITAL COPIES WILL BE PROVIDED.

**PRODUCT INFORMATION**

**Item:** APL-GLIDER.LEGACY  
**Item Description:** (LEGACY) APL Glider  
**Serial:** 0338

**Special Notes**

Services Requested:  
 Conductivity Calibration  
 Temperature Calibration

Services Performed:  
 Perform initial diagnostic evaluation.  
 Performed pressure calibration.  
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_SEAGLIDER	Calibrate Seaglider conductivity and temperature sensors	1

**Unbilled Items**

Item	Item Description	Qty



**Sea-Bird Electronics, Inc.**  
 13431 NE 20th Street  
 Bellevue, WA 98005 United States

Phone  
 Fax

+1-425-643-9866  
 +1-425-643-9954  
 www.seabird.com

**SERVICE REPORT**

**Service Request**  
**Date**  
**Sales Order**

**1005510402**  
 22-MAY-2020  
 316003277

**PRODUCT INFORMATION**

**Item:** APL-GLIDER.LEGACY  
**Item Description:** (LEGACY) APL Glider  
**Serial:** 0351

**Special Notes**

Services Requested:  
 Conductivity Calibration  
 Temperature Calibration

Problems Found:

Temperature found to have drifted; replacement recommended.

Services Performed:

Perform initial diagnostic evaluation.  
 Performed pressure calibration.  
 Performed "POST" cruise calibration.  
 Replaced the temperature probe.  
 Performed a "Final" Calibration.

Item	Item Description	Qty
CAL_SEAGLIDER	Calibrate Seaglider conductivity and temperature sensors	1

**Unbilled Items**

Item	Item Description	Qty
802134	VM GLIDER THERM PROBE ASSY,V3,SHIELDED,FENWAL	1



**Sea-Bird Electronics, Inc.**  
 13431 NE 20th Street  
 Bellevue, WA 98005 United States

Phone  
 Fax

+1-425-643-9866  
 +1-425-643-9954  
 www.seabird.com

**SERVICE REPORT**

**Service Request**  
**Date**  
**Sales Order**

**1005510402**  
 22-MAY-2020  
 316003277

**PRODUCT INFORMATION**

**Item:** APL-GLIDER.LEGACY  
**Item Description:** (LEGACY) APL Glider  
**Serial:** 0309

**Special Notes**

Services Requested:  
 Conductivity Calibration  
 Temperature Calibration

Problems Found:  
 No problems found.

Services Performed:  
 Perform initial diagnostic evaluation.  
 Performed pressure calibration.  
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_SEAGLIDER	Calibrate Seaglider conductivity and temperature sensors	1

**Unbilled Items**

Item	Item Description	Qty



**Sea-Bird Electronics, Inc.**  
 13431 NE 20th Street  
 Bellevue, WA 98005 United States

Phone  
 Fax

+1-425-643-9866  
 +1-425-643-9954  
 www.seabird.com

**SERVICE REPORT**

**Service Request**  
**Date**  
**Sales Order**

**1005510402**  
 22-MAY-2020  
 316003277

**PRODUCT INFORMATION**

**Item:** APL-GLIDER.LEGACY  
**Item Description:** (LEGACY) APL Glider  
**Serial:** 0352

**Special Notes**

Services Requested:  
 Conductivity Calibration  
 Temperature Calibration

Problems Found:  
 High positive conductivity drift; replacement recommended.

Services Performed:  
 Perform initial diagnostic evaluation.  
 Performed pressure calibration.  
 Performed "POST" cruise calibration.  
 Replaced the conductivity cell.  
 Performed a "Final" Calibration.

Item	Item Description	Qty
CAL_SEAGLIDER	Calibrate Seaglider conductivity and temperature sensors	1

**Unbilled Items**

Item	Item Description	Qty
802761	ALACE COND CELL ASSEMBLY, NO WRAP, SHIELDED WIRES	1