

13431 NE 20th Street Bellevue, WA 98005 United States Phone +1-425-643-9866

www.seabird.com

1005516731

Service Request Date 03-FEB-2023 Sales Order 320430494

SERVICE REPORT

CUSTOMER INFORMATION

Name: NOAA - AOML ATLANTIC OCEANOGRAPHIC and METEOROLOGICAL

LAB

Account: 40279759 **GRANT RAWSON** GRANT.RAWSON@NOAA.GOV (305) 361-4363

Bill To Address

NOAA - AOML ATLANTIC OCEANOGRAPHIC and METEOROLOGICAL LAB 4301 RICKENBACKER CAUSEWAY MIAMI,FL,33149,US

Ship To Address

NOAA - AOML ATLANTIC OCEANOGRAPHIC and METEOROLOGICAL LAB 4301 RICKENBACKER CAUSEWAY MIAMI,FL,33149,US

PO Number:

PRODUCT INFORMATION

Item: APL-GLIDER.LEGACY

Item Description: (LEGACY) APL Glider

Serial: 0316

Special Notes

Services Requested: CAL_SEAGLIDER **CNCRTSEAGLDR**

Problems Found:

The temperature was found to have excessive drift.

Services Performed:

Perform initial diagnostic evaluation. Performed "POST" cruise calibration. Replaced the temperature probe. Replaced the O-rings.

Performed a "Final" Calibration.

| Item | Item Description | Qty |
|---------------|--|-----|
| CAL_SEAGLIDER | Calibrate Seaglider conductivity and temperature sensors | 1 |
| CNCRTSEAGLDR | CONFIRM & RE-CERTIFY SEAGLIDER CTD | 1 |
| REPGLIDERPROB | Replace SBE Glider thermistor probe | 1 |
| | | |

| Item | Item Description | Qty |
|--------|--|-----|
| 802134 | VM GLIDER THERM PROBE ASSY, V3, SHIELDED, FENWAL | 1 |
| | | |



SERVICE REPORT

Sea-Bird Electronics, Inc.

13431 NE 20th Street Bellevue, WA 98005 United States Phone +1-425-643-9866

www.seabird.com

Service Request Date Sales Order **1005516731** 03-FEB-2023 320430494

PRODUCT INFORMATION

Item: APL-GLIDER.LEGACY

Item Description: (LEGACY) APL Glider

Serial: 0366

Special NotesProblems Found:

The temperature was found to have excessive drift.

Services Performed:
Perform initial diagnostic evaluation.
Performed "POST" cruise calibration.
Replaced the temperature probe.
Replaced the O-rings.

Performed a "Final" Calibration.

| Item | Item Description | Qty |
|---------------|--|-----|
| CAL_SEAGLIDER | Calibrate Seaglider conductivity and temperature sensors | 1 |
| CNCRTSEAGLDR | CONFIRM & RE-CERTIFY SEAGLIDER CTD | 1 |
| REPGLIDERPROB | Replace SBE Glider thermistor probe | 1 |

| Item | Item Description | Qty |
|--------|---|-----|
| 802134 | VM GLIDER THERM PROBE ASSY,V3,SHIELDED,FENWAL | 1 |
| | | |



13431 NE 20th Street Bellevue, WA 98005 United States Phone +1-425-643-9866

www.seabird.com 1005516731

03-FEB-2023 320430494

SERVICE REPORT

Service Request Date Sales Order

PRODUCT INFORMATION

Item: APL-GLIDER.LEGACY

Item Description: (LEGACY) APL Glider

Serial: 0360

Special Notes
Services Performed:
Perform initial diagnostic

Perform initial diagnostic evaluation. Performed "POST" cruise calibration.

| Item | Item Description | Qty |
|---------------|--|-----|
| CAL_SEAGLIDER | Calibrate Seaglider conductivity and temperature sensors | 1 |
| CNCRTSEAGLDR | CONFIRM & RE-CERTIFY SEAGLIDER CTD | 1 |

| Item | Item Description | Qty |
|------|------------------|-----|
| | | |



13431 NE 20th Street Bellevue, WA 98005 United States Phone +1-425-643-9866

Date

Sales Order

www.seabird.com

Service Request 1005516731 03-FEB-2023 320430494

SERVICE REPORT

PRODUCT INFORMATION

Item: APL-GLIDER.LEGACY

Item Description: (LEGACY) APL Glider

Serial: 0346

Special Notes Services Performed:

Perform initial diagnostic evaluation. Performed "POST" cruise calibration.

| Item | Item Description | Qty |
|---------------|--|-----|
| CAL_SEAGLIDER | Calibrate Seaglider conductivity and temperature sensors | 1 |
| CNCRTSEAGLDR | CONFIRM & RE-CERTIFY SEAGLIDER CTD | 1 |

| | Item | Item Description | Qty |
|--|------|------------------|-----|
| | | | |



SERVICE REPORT

Sea-Bird Electronics, Inc.

13431 NE 20th Street Bellevue, WA 98005 United States Phone +1-425-643-9866

www.seabird.com

Service Request Date Sales Order

1005516731 03-FEB-2023 320430494

Item: APL-GLIDER.LEGACY

PRODUCT INFORMATION

Item Description: (LEGACY) APL Glider

Serial: 0347

Special Notes

Services Performed:

Perform initial diagnostic evaluation. Performed "POST" cruise calibration. Replaced the temperature probe. * Replaced the O-rings.

Performed a "Final" Calibration.

^{*} Probe was found to have excessive drift after order was original quoted for repairs. This was done at no charge.

| Item | Item Description | Qty |
|---------------|--|-----|
| CAL_SEAGLIDER | Calibrate Seaglider conductivity and temperature sensors | 1 |
| CNCRTSEAGLDR | CONFIRM & RE-CERTIFY SEAGLIDER CTD | 1 |
| | | |

| Item | Item Description | Qty |
|--------|---|-----|
| 802134 | VM GLIDER THERM PROBE ASSY,V3,SHIELDED,FENWAL | 1 |
| | | |



13431 NE 20th Street Bellevue, WA 98005 United States Phone +1-425-643-9866

www.seabird.com

Service Request Date Sales Order 1005516731 03-FEB-2023 320430494

SERVICE REPORT

PRODUCT INFORMATION

Item: APL-GLIDER.LEGACY

Item Description: (LEGACY) APL Glider

Serial: 0378

Special Notes
Services Performed:

Perform initial diagnostic evaluation. Performed "POST" cruise calibration.

| Item | Item Description | Qty |
|---------------|--|-----|
| CAL_SEAGLIDER | Calibrate Seaglider conductivity and temperature sensors | 1 |
| CNCRTSEAGLDR | CONFIRM & RE-CERTIFY SEAGLIDER CTD | 1 |

| Item | Item Description | Qty |
|------|------------------|-----|
| | | |