



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005511990
 14-DEC-2020
 316236648

CUSTOMER INFORMATION

Name: NOAA - ATLANTIC OCEANOGRAPHIC AND METEOROLOGICAL LAB
 Account : 40279759

PO Number:
 EA133M-15-CQ-0039 /
 1333MJ20FNRMA0011

GRANT RAWSON
 GRANT.RAWSON@NOAA.GOV
 (305) 361-4363

Bill To Address
 4301 RICKENBACKER CAUSEWAY
 MIAMI,FL,33149,US

Ship To Address
 4301 RICKENBACKER CAUSEWAY
 MIAMI,FL,33149,US

CUSTOMER NOTES:

1333MJ20FNRMA0011 \$20K TASK ORDER AGAINST IDIQ EA133M-15-CQ-0039 for payment and FedEx act 249391433 for standard overnight return shipping.

REQUESTED SERVICE DETAILS:
 UPDATE FIRMWARE: Yes

REQUESTED TRANSACTION DETAILS:

PO NAME: See notes for task order #
 PAPER COPIES: Yes
 REQUESTED RETURN: 05-Feb-2021
 SPLIT SHIP: split

PRODUCT INFORMATION

Item: 43.LEGACY
Item Description: (LEGACY) SBE 43 Oxygen Sensor
Serial: 432082

Special Notes

Services Requested:
 Evaluate/Repair Instrumentation.
 Perform Routine Calibration Service.

Services Performed:

Perform initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005511990
 14-DEC-2020
 316236648

PRODUCT INFORMATION

Item: 43.LEGACY
Item Description: (LEGACY) SBE 43 Oxygen Sensor
Serial: 43-2940

Special Notes

Services Requested:
 Evaluate/Repair Instrumentation.
 Perform Routine Calibration Service.

Services Performed:
 Perform initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005511990
 14-DEC-2020
 316236648

PRODUCT INFORMATION

Item: 43.LEGACY
Item Description: (LEGACY) SBE 43 Oxygen Sensor
Serial: 432083

Special Notes

Services Requested:
 Evaluate/Repair Instrumentation.
 Perform Routine Calibration Service.

Services Performed:
 Perform initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005511990
 14-DEC-2020
 316236648

PRODUCT INFORMATION

Item: 04C.210
Item Description: SBE4C COND SENSOR V2, 6800M, XSG, SCOTCHCAST
Serial: 04-4709

Special Notes

Services Requested:
 Conductivity Calibration

Services Performed:
 Perform initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005511990
 14-DEC-2020
 316236648

PRODUCT INFORMATION

Item: 04C.210
Item Description: SBE4C COND SENSOR V2, 6800M, XSG, SCOTCHCAST
Serial: 04-4349

Special Notes

Services Requested:
 Conductivity Calibration

Services Performed:
 Perform initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005511990
 14-DEC-2020
 316236648

PRODUCT INFORMATION

Item: 04C.210
Item Description: SBE4C COND SENSOR V2, 6800M, XSG, SCOTCHCAST
Serial: 04-4346

Special Notes

Services Requested:
 Conductivity Calibration

Services Performed:
 Perform initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005511990
 14-DEC-2020
 316236648

PRODUCT INFORMATION

Item: 03P.210
Item Description: SBE3PLUS TEMP, GE PROBE, 6800M, XSG
Serial: 03P5889

Special Notes

Services Requested:
 Temperature Calibration

Services Performed:
 Perform initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_3	SBE 3 Temperature sensor calibration; 11 points -1 to 31deg.C	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005511990
 14-DEC-2020
 316236648

PRODUCT INFORMATION

Item: 03F.210
Item Description: SBE3F FAST TEMP, GE PROBE, 6800M, XSG
Serial: 03P5855

Special Notes

Services Requested:
 Temperature Calibration

Services Performed:
 Perform initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_3	SBE 3 Temperature sensor calibration; 11 points -1 to 31deg.C	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005511990
 14-DEC-2020
 316236648

PRODUCT INFORMATION

Item: 03P.210
Item Description: SBE3PLUS TEMP, GE PROBE, 6800M, XSG
Serial: 03P5898

Special Notes

Services Requested:
 Temperature Calibration

Services Performed:
 Perform initial diagnostic evaluation.
 Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_3	SBE 3 Temperature sensor calibration; 11 points -1 to 31deg.C	1

Unbilled Items

Item	Item Description	Qty