

Cruise: EX0924
Ship: M/V Explorer of the Seas
Dates: July 27 - 31, 2009
Expocode: 33KF20090723
Chief Scientist: not applicable
Equipment: Surface samples collected.
Total number of stations: VOS Underway Cruise

Sample Collection

The discrete samples were collected by Kevin Sullivan from a metering ball valve next to the underway pCO₂ instrument. The underway pCO₂ instrument was located in the bow thruster space next to the TSG and a short distance from the inlet pump. The sea water takes less than 10 seconds to travel from the inlet to the instruments. The TSG temperature is believed to be no more than 0.15 degrees C warmer than in-situ SST. The date and time listed in the data file are UTC when each sample bottle was collected.

DIC:

9 locations, 18 samples each 500-ml, 9 sets of duplicate samples
Sample_ID#: 501 - 518
PI: Dr. Rik Wanninkhof
Analyzed by: Esa Peltola

Talk:

9 locations, 18 samples each 500-ml, 9 sets of duplicate samples
Sample_ID#: 501 - 518
PI: Dr. John Morse
Analyzed by: Luz Romero

Sample Analysis

DIC:

Analysis date: September 14, 2009
Coulometer used: AOML2
Blank: 14.7-25 counts/min
CRM # used and assigned value (include both DIC and salinity): Batch 85, c: 2000.4 umol/kg,S: 33.326
CRM value measured: AOML 2: offset 5.9 umol/kg (2006.3 umol/kg)
Average run time, minimum run time, maximum run time: 15 min, 10 min, 20 min
Reproducibility: (# samples and average difference): 9 sets of duplicate samples, average difference 0.8 umol/kg
CRM, salinity and HgCl₂ correction applied: Salinity correction was applied using TSG salinity
Remarks-

The volume correction was applied due to added HgCl₂ (Measured DIC*1.00037).
The first CRM of each cell was used for a CRM correction.
The computer stalled after sample run 517, but this had no effect on the sample runs.

Talk:

The results posted are duplicate analyses from the same sample bottle.

Comments

The GPS transducer and the thermosalinograph (TSG) were logged by a computer system installed by NOAA.

The latitude, longitude, temperature and salinity reported with the DIC and TALK measurements were taken from the raw TSG data file. The merging of the discrete measurements with the TSG data was done on the basis of date and time. The TSG values are provided for reference; no post-cruise assurance of accuracy has been done to this data.

The Sample_ID is the sample bottle number for the discrete samples.

UPDATE:

Between March and June of 2021, all of the data for the discrete samples was put into a uniform format. The supporting information was checked for accuracy, especially the expocode, date, time, and positions.