

**Cruise:** EQNX\_20160415  
**Ship:** Royal Caribbean Equinox  
**Expocode:** MLCE20160415  
**Dates:** April 15<sup>th</sup> – April 25<sup>th</sup>, 2016  
**Chief Scientist:** Dr. Denis Pierrot  
**Equipment:** TSG-Flow thru system  
**Total number of stations:** 36  
**Location:** Ft. Lauderdale, FL to Lisbon, Portugal

### **Sample Collection**

The discrete samples were collected from the TSG-flow thru system onboard the ship of opportunity Royal Caribbean Equinox by Dr. Denis Pierrot. The date and time listed in the data file are UTC when each sample bottle was collected.

#### **DIC:**

36 locations, 40 samples each 500-ml, 4 duplicate samples.  
Sample\_ID#: 301, etc.; Sample bottle number  
PI: Dr. Rik Wanninkhof  
Analyzed by: Charles Featherstone

#### **pH:**

36 locations, 40 samples each 500-ml, 4 duplicate samples.  
Sample\_ID#: 301, etc.; Sample bottle number  
PI: Dr. Rik Wanninkhof  
Analyzed by: Charles Featherstone

#### **TALK:**

36 locations, 40 samples each 500-ml, 4 duplicate samples.  
Sample\_ID#: 301, etc.; Sample bottle number  
PI: Dr. Rik Wanninkhof  
Analyzed by: Dr. Leticia Barbero and Charles Featherstone

### **Sample Analysis**

#### **DIC:**

Instrument ID	Date	Certified CRM (µmol/kg)	CRM Value (µmol/kg)	CRM Offset (µmol/kg)	Blank (Counts)	Avg. Sample Analysis Time
AOML 3	07/18/2016	2031.53	2031.26	0.27	26.0	15
AOML 3	07/19/2016	2031.53	2033.04	1.51	26.0	14
AOML 4	07/18/2016	2031.53	2028.56	2.97	28.0	13

Analysis date: 07/18/2016  
 Coulometer used: DICE–CM5015- AOML 3  
 Blanks: 14.4 counts/min and raised to 26.0 counts/min before CRM analysis  
 CRM # 1235 was used and with an assigned value of (includes both DIC and salinity):  
 Batch 144, c: 2031.53  $\mu\text{mol/kg}$ , S: 33.571  
 CRM values measured: AOML 3: offset 0.27  $\mu\text{mol/kg}$  (2031.26  $\mu\text{mol/kg}$ ).  
 Average run time, minimum run time, maximum run time: 15, 12 and 19 min.

Analysis date: 07/19/2016  
 Coulometer used: DICE–CM5015- AOML 3  
 Blanks: 12.0 counts/min and raised to 26.0 counts/min before CRM analysis  
 CRM # 1110 was used and with an assigned value of (includes both DIC and salinity):  
 Batch 144, c: 2031.53  $\mu\text{mol/kg}$ , S: 33.571  
 CRM values measured: AOML 3: offset 1.51  $\mu\text{mol/kg}$  (2033.04  $\mu\text{mol/kg}$ ).  
 Average run time, minimum run time, maximum run time: 14, 11 and 16 min.

Analysis date: 07/19/2016  
 Coulometer used: DICE–CM5015- AOML 4  
 Blanks: 23.9 counts/min and raised to 28.0 counts/min before CRM analysis  
 CRM # 236 was used and with an assigned value of (includes both DIC and salinity):  
 Batch 144, c: 2031.53  $\mu\text{mol/kg}$ , S: 33.571  
 CRM values measured: AOML 4: offset 2.97  $\mu\text{mol/kg}$  (2028.56  $\mu\text{mol/kg}$ ).  
 Average run time, minimum run time, maximum run time: 13, 12 and 16 min.

**Reproducibility:** (# samples and average difference): 4 sets of duplicate samples, average difference 5.00  $\mu\text{mol/kg}$  (3.62-6.03), average STDEV of 3.54 (2.56-4.26).

Instrument ID	Sample ID	Bottle #	Corrected DIC ( $\mu\text{mol/kg}$ )	Average	Difference	STDEV
AOML3	10000	1	2036.64			
AOML4	20000	2	2042.53	2039.59	5.88	4.16
AOML4	80000	8	2081.93			
AOML3	90000	9	2085.55	2083.74	3.62	2.56
AOML3	260000	26	2115.68			
AOML3	270000	27	2109.65	2112.66	6.03	4.26
AOML3	820000	82	2123.00			
AOML3	830000	83	2118.53	2120.76	4.47	3.16
Average					5.00	3.54

CRM, salinity and HgCl<sub>2</sub> correction applied: Salinity correction was applied using TSG salinity.

## Remarks

The volume correction was applied due to added HgCl<sub>2</sub> (Measured DIC\*1.00037).  
The first CRM of each cell was used for a CRM correction.

The DIC instruments were stable: the gas loop and CRM values did not change significantly throughout the life span of each cell.

The samples were analyzed using the DICE (AOML 3 & AOML 4) and a new coulometer from UIC, Inc. CM5015 with CM5011 emulation software.

Duplicates were sampled 2 to 4 minutes apart from the ships TSG flow thru system.

## **pH:**

Analysis date: 07/18/2016 and 07/19/2016  
Spectrophotometer used: HP Agilent 8453

**Reproducibility:** (# samples and average difference): 4 sets of duplicate samples, average difference 0.0061 (0.0019-0.0125), average STDEV of 0.0043 (0.0013-0.0088).

System	Sample ID	Bottle #	pH	Average	Difference	STDEV
HP Agilent 8453	10000	1	8.1387			
HP Agilent 8453	20000	2	8.1405	8.1396	0.0019	0.0013
HP Agilent 8453	80000	8	8.1183			
HP Agilent 8453	90000	9	8.1153	8.1168	0.0030	0.0021
HP Agilent 8453	260000	26	8.0382			
HP Agilent 8453	270000	27	8.0312	8.0347	0.0071	0.0050
HP Agilent 8453	820000	82	8.0238			
HP Agilent 8453	830000	83	8.0114	8.0176	0.0125	0.0088
Average					0.0061	0.0043

## Temperature measured during pH analysis

Sample ID	Sample BTL #	Btl. Temp	Start Cell (°C)	End Cell (°C)	Differ Start to End Cell (°C)
10000	1	19.821	20.918	21.648	0.730
20000	2	19.802	20.712	21.509	0.797
30000	3	20.004	20.907	21.694	0.787
40000	4	20.009	20.949	21.862	0.913
50000	5	19.986	20.789	21.911	1.122
60000	6	20.303	20.965	21.582	0.617

70000	7	20.070	20.898	21.798	0.900
80000	8	19.958	20.927	21.926	0.999
90000	9	19.974	20.886	21.735	0.849
100000	10	19.951	20.828	21.939	1.111
110000	11	20.004	20.855	21.962	1.107
120000	12	20.068	20.901	21.736	0.835
130000	13	20.126	20.952	21.618	0.666
140000	14	20.012	20.961	21.952	0.991
160000	16	20.152	20.951	21.804	0.853
200000	17	19.960	20.932	21.922	0.990
170000	18	19.985	20.865	21.810	0.945
220000	19	20.029	20.986	21.880	0.894
190000	20	19.916	20.849	21.703	0.854
210000	21	19.964	20.805	21.714	0.909
230000	22	19.954	20.982	21.972	0.990
250000	23	20.003	20.987	21.933	0.946
270000	24	20.002	20.892	21.919	1.027
310000	25	20.024	20.838	21.796	0.958
240000	26	20.037	20.838	21.916	1.078
260000	27	19.957	20.808	21.779	0.971
290000	29	19.949	20.904	21.753	0.849
350000	31	20.109	21.076	21.963	0.887
370000	33	20.043	20.947	21.674	0.727
390000	35	20.07	21.054	21.927	0.873
330000	37	19.863	20.815	21.699	0.884
810000	39	20.131	20.962	21.896	0.934
820000	81	19.897	20.877	21.594	0.717
830000	82	18.896	20.695	21.680	0.985
880000	83	20.024	20.935	21.723	0.788
890000	88	20.045	20.957	21.826	0.869
900000	89	19.956	20.822	21.630	0.808
910000	90	20.044	20.862	21.655	0.793
1740000	91	19.987	20.818	21.740	0.922
180000	174	19.984	20.708	21.686	0.978
Average		19.977	20.890	21.787	0.896

### **Remarks**

Duplicates were sampled 2 to 4 minutes apart from the ships TSG flow thru system.

The equations of Liu et al, 2011 formulated using the purified m-cresol purple indicator was used to determine pH of the samples. pH samples were analyzed at 20<sup>0</sup>C at Full Scale (pH 0-14).

Temperature for each sample was measured before analysis using a Hart Scientific Fluke 1523 reference thermometer.

Approximately 80 mL of sample was extracted from each DIC sample bottle by syringe before DIC analysis to determine the pH.

**Talk:**

Analysis date: 7/28/2016

Titration system used: Open cell

CRM Batch 120, Salinity = 33.072, cert. TA = 2208.34  $\mu\text{mol/kg}$ .

On 7/28/2016 one CRM was analyzed before the samples and another CRM was run at the end of analysis for each system.

The TA for the water samples was corrected using the daily averaged ratios between the certified and measured values of the CRMs run on each cell. The following table shows the CRM measurements for each day and cell.

Cell System	Date	Time	Bottle #	TA	$ \Delta\text{CRM} $
1	07/28/2016	11:26:45	434	2207.70	
1	07/28/2016	19:57:20	434	2211.51	3.81
2	07/28/2016	11:19:11	268	2182.56	
2	07/28/2016	19:50:56	268	2183.19	0.63

**Reproducibility:** (# samples and average difference): 4 duplicate samples were collected with an average difference 3.78  $\mu\text{mol/kg}$  (0.34 – 9.89) and an average STDEV of 2.67 (0.24 – 6.99).

System	Sample ID	Talk	Average	Difference	STDEV
System 2	10000	2368.62	2373.56	9.89	6.99
System 1	20000	2378.51			
System 1	80000	2400.61	2401.45	1.68	1.19
System 2	90000	2402.29			
System 1	260000	2377.77	2376.16	3.21	2.27
System 2	270000	2374.56			
System 2	820000	2365.48	2365.65	0.34	0.24
System 1	830000	2365.83			

---

Average	3.78	2.67
---------	------	------

---

### **Remarks**

The CRM measurement for each day was used to correct the data for that day only. Both systems worked well.

Duplicates were sampled 2 to 4 minutes apart from the ships TSG flow thru system.

### **Comments**

The latitude, longitude, date, and time reported with the DIC, pH and TALK measurements were taken from the sample field log. The field log values are provided for reference; no post-cruise assurance of accuracy has been done to this data.

The Sample ID is the sample station, cast number and Niskin bottle number for the discrete samples.

The salinity and temperature were taken from the UW pCO<sub>2</sub> system. The salinities were used in the DIC, Talk and pH calculations.

Corresponding UW pCO<sub>2</sub> data can be found at the following website  
<http://www.aoml.noaa.gov/ocd/ocdweb/occ.html>

### UPDATE:

Between March and May of 2021, all of the data for the discrete samples was put into a uniform format. The supporting information was checked for accuracy, especially the expocode, date, time, and positions.

The ship code in the Expocode was corrected (EQNX => MLCE).

Additionally, pH results were recalculated to 20 and 25 degrees Celsius.